



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L2904-230905)

This is to certify that

Wei-I Industry Co., Ltd

WEI-I Inspection Center

381, ShunAn Rd., Luzhu District, Kaohsiung City 82150, Taiwan

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 2904

Originally Accredited : April 17, 2014

Effective Period : September 05, 2023 to September 04, 2026

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Ching-Chang Lien

Ching-Chang Lien
President, Taiwan Accreditation Foundation
September 05, 2023

Accreditation Number : 2904

Laboratory Head : LIU, Yi-Xin

01. 03 Metals and Alloys Products

Nuts

M014 Proof Load (Compression Method)

ASTM F606/F606M

ISO 898-2

SAE J995

(4.9 to 441) kN

(500 to 45000) kgf

Approval Signatory: LIU, Yi-Xin

M017 Internal Thread Parameters-System 21

ASME B1.2

ISO 1502

M3 to M30

#6, (1/4, 3/8, 5/8, 3/4, 7/8, 1, 1-1/8, 1-5/16, 1-7/16) in

Approval Signatory: LIU, Yi-Xin

M017 Dimensions of General Purpose Fasteners and High-Volume Machine Assembly Fasteners

JIS B1071

ISO 4759-1

Caliper: (0 to 200) mm

Outside Micrometer: (0 to 50) mm

Height Gage: (0 to 50) mm

Video Measurement System: X-Axis: (0 to 50) mm; Y-Axis: (0 to 50) mm; Radius: (0 to 12.7) mm; Angle: (0 to 90) °

Approval Signatory: LIU, Yi-Xin

M101 Rockwell Hardness

ASTM E18

ISO 898-2

ASTM F606/F606M

(40 to 100) HRBW

(20 to 65) HRC

Approval Signatory: LIU, Yi-Xin

M153 Surface Discontinuities

ASTM F812 (Exclude The Cone Proof Load Test)

ISO 6157-2 (Exclude Destructive Testing And The Cone Proof Load Test)

(1/4 to 1 1/2) in

M5 to M39

Approval Signatory: LIU, Yi-Xin

(Null below)

P2, total 2 pages

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix

